

Certificate



for European Product Safety

Reference No. 1309927

Medical Power Supply

Type designation GT*41078-*05-USB

Certificate holder **GlobTek, Inc.**
186 Veterans Dr. Northvale, NJ 07647,
USA

The product complies with the standard(s) EN 60601-1: 2006 + A11: 2011

Date of expiry 01 August 2018

EU Directive information The product satisfies the provisions for CE marking according to the Low Voltage Directive 2006/95/EC.

Additional information in Appendix.

Certification Body Intertek Semko AB, Product Certification

Place Shanghai

Signed



Gary Hu

Date 01 August 2013
Page 1 of 2

This certificate is issued in accordance with the terms and conditions set out in the Appendix.

**Intertek**
Appendix

Reference No. 1309927

*Technical data**Type designation*

GT*41078-*05-USB

Input

100-240V~, 50-60Hz, 0.3A, Class II

BRAND name:

GlobTek

Product information

The 1st "*" part can be 'M' or '-' or 'H' for market identification and not related to safety.

The 2nd "*" part denotes the rated output wattage designation, which can be "01" to "06", with interval of 1.

Test performed on model GTM41078-0605-USB as worst case, and also performed on other output models for reference. Although ME equipment installation and use type are not specified, mechanical strength tests including push test, 1 m height drop test and mould stress relief test are performed considering the usage for adapter model.

Manufacturing site(s)

GlobTek (Suzhou) Co., Ltd
Building 4, No. 76, Jin Ling East Rd., Suzhou Industrial Park, Suzhou,
JiangSu, 215021, P.R.China

Safety related parts of the user instructions shall be written in a language acceptable to the country where the product is to be used.

This certificate is based on a CB certificate with ref. No. SE-73537 dated 16 Jul 2013, issued by Intertek Semko Ab.

This certificate, which includes production control, constitutes sufficient support for an EC Declaration of Conformity and CE marking, according the principle of presumption of conformity, of the product according to the Low Voltage Directive 2006/95/EC.

This presumption can expire before end of validity of this certificate due to new issued Standard or Amendment and changes within the EU legislation.

Certification Body

Intertek Semko AB, Product Certification

Place Shanghai*Signed*

Gary Hu

Date 01 August 2013

This Certificate is for the exclusive use of Intertek's Client and is provided pursuant to the Certification agreement between Intertek and its Client. Intertek's responsibility and liability are limited to the terms and conditions of the agreement. Intertek assumes no liability to any party, other than to the Client in accordance with the agreement, for any loss, expense or damage occasioned by the use of this certificate. Only the Client is authorized to permit copying or distribution of this certificate and then only in its entirety. Use of Intertek's Certification mark is restricted to the conditions laid out in the agreement. Any further use of the Intertek name for the sale or advertisement of the tested material, product or service must first be approved in writing by Intertek. Initial Factory Assessments and Follow up Services are for the purpose of assuring appropriate usage of the Certification mark in accordance with the agreement, they are not for the purposes of production quality control and do not relieve the Client of their obligations in this respect.