



Ref. Certif. No.

SE-87034

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Medical Power Supply

Name and address of the applicant

GlobTek, Inc.
186 Veterans Dr. Northvale, NJ 07647 USA

Name and address of the manufacturer

Same as applicant

Name and address of the factory


Note: When more than one factory, please report on page 2

See page 2

Ratings and principal characteristics

See page 2

Trademark (if any)



Customer's Testing Facility (CTF) Stage used

-

Model / Type Ref.

GT*961200P****, GT*96900P****, GT*41133-****

Additional information (if necessary may also be reported on page 2)

See page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005+A1
IEC 60601-1-11:2015

As shown in the Test Report Ref. No. which forms part of this Certificate

161200816SHA-001
161200916SHA-002

This CB Test Certificate is issued by the National Certification Body

Intertek Semko AB
Box 1103
SE-164 22 Kista, Sweden
Int +46 8 750 00 00

Date: 18 May 2017

Intertek

Signature:

Gary Mu

Factories:

GlobTek (Suzhou) Co., Ltd
Building 4, No. 76, Jin Ling East Rd., Suzhou Industrial Park, Suzhou, JiangSu 215021, China

GlobTek, Inc.
186 Veterans Dr. Northvale, NJ 07647 USA

Ratings and principal characteristics:

GT*961200P****, GT*96900P****:

Input: 100-240V~, 50-60Hz, 1.5A;

Output: 12-54Vdc, max. 120W or 90W

GT*41133-****:

Input: 100-240V~, 50-60Hz or 50-400Hz, 1.5A;

Output: 12-48Vdc, max. 90W

Additional information (if necessary)

Explanation of type designation GT*961200P**** and GT*96900P**** and GT*41133-****

The 1st symbol "*" can be 'M' or '.' or 'H' for market identification and not related to safety.

When model = GT*41133-****

The 2nd symbol "*" denotes the rated output wattage designation, which can be "01" to "90", with interval of 1.

The 3rd symbol "*" denotes the standard rated output voltage designation, which can be "16", "24", "35" and "48".

The 4th symbol "*" is optional, which can be "-0.1" to "-12.9" with interval of 0.1 to denote voltage deviation or blank to indicate no voltage different.

The 3rd symbol "*" and 4th symbol "*" together denote the output voltage, with a range of 12 - 48 volts

The 5th symbol "*" can be

- T2 means desktop class II with C8 AC inlet
- T3A means desktop class I with C6 AC inlet
- F means Open Frame class I
- FW means Open Frame class II

The last symbol "*" denote any six character = 0-9 or A-Z or ()[] or - or blank for marketing purposes.

Tests were performed on 12Vdc/7.5A, 37.5Vdc/2.4A and 48Vdc/1.875A output power adapter model as representative, and also performed on other output models for reference. The clearance & creepage distance measurement, mechanical strength and temperature rising of open frame model shall be reevaluated in end product combined with this report.

Each standard rated output voltage designation corresponds to a transformer model. Each transformer model is identical in insulation construction including clearance and creepage except number of turns per coil.

When model = GT*961200P**** and GT*96900P****

The 2nd symbol "*" denotes the rated output wattage designation, which can be "-01" to "-120", with interval of 1 and "-" can be omitted.

The 3rd symbol "*" denote the standard rated output voltage designation, which can be "12" to "54" or "12.0" to "54.0" in 0.1V increments

The 4th symbol "*" can be

- T2 means desktop class II with C8 AC inlet
- T2A means desktop class II with C18 AC inlet
- T3 means desktop class I with C14 AC inlet
- T3TAB means desktop class I with C14 AC inlet and housing with a tab.
- T3A means desktop class I with C6 AC inlet

The last symbol "*" denote any six character = 0-9 or A-Z or ()[] or - or blank for marketing purposes

Date: 18 May 2017

Signature: 