

## Certificate



Reference No. 1917153

## Medical Power supply

Type designation

GT\*961200P\*\*\*\*, GT\*96900P\*\*\*\*, GT\*41133-\*\*\*\*

Certificate holder

GlobTek, Inc.

186 Veterans Dr. Northvale, NJ 07647

USA

The product complies with

the standard(s)

EN 60601-1:2006+A1+A12

EN 60601-1-11:2015

Date of expiry

10 July 2022

Additional information in Appendix.

Certification Body

Intertel, Semko AB, Product Certification

Place Kista - Stockholm

Signed

Henrik Wikström

Date

14 October 2019
Page 1 of 4

This certificate is issued in accordance with the terms and conditions set out in the Appendix.





Reference No. 1917153

Input:100-240V~, 50-60Hz or 50-400Hz, 1.5A

Technical data

Type designation Input

For GT\*961200P\*\*\*\*, GT\*96900P\*\*\*\*, GT\*41133-\*\*\*\*\*

For GT\*961200P\*\*\*\* and GT\*96900P\*\*\*\*:

Input:100-240V~,50-60Hz, 1.5A

For GT\*41133-\*\*\*\*:

Output BRAND name:

See the attachment

Product information:

Model Similarity:

GT\*961200P\*\*\*\* and GT\*96900P\*\*\*\* and GT\*41133-\*\*\*\*\*

The 1st "\*" part can be 'M' or '-' or 'H' for market identification and not related to safety.

When model = GT\*41133-\*\*\*\*:

The 2nd "\*" denotes the rated output wattage designation, which can be "01" to "90", with interval of 1. The 3rd "\*" denotes the standard rated output voltage designation, which can be "16", "24", "35" and "48". The 4th "\*" part is optional, which can be "-0.1" to "-12.9" with interval of 0.1 to denote voltage deviation or blank to indicate no voltage different.

The 3rd "\*" and 4th "\*" together denote the output voltage, with a range of 12 - 48 volts. The 5th "\*"

=-T2 means desktop class II with C8 AC inlet

- =-T3A means desktop Class I with C6 AC inlet
- =-F means Open Frame Class I
- =-FW means Open Frame class II

The last \* denote any six character = 0-9 or A-Z or ()[] or – or blank for marketing purposes.

Tests were performed on 12Vdc/7.5A, 37.5Vdc/2.4A and 48Vdc/1.875A output power adapter model as representative, and also performed on other output models for reference. The clearance & creepage distance measurement, mechanical strength and temperature rising of open frame model shall be revaluated in end product combined with this report.

Each standard rated output voltage designation corresponds to a transformer model. Each transformer model is identical in insulation construction including clearance and creepage except number of turns per coil.

When model = GT\*961200P\*\*\*\* and GT\*96900P\*\*\*\*

The 2nd "\*" denotes the rated output wattage designation, which can be "-01" to "-120", with interval of 1 and "-" can be omitted.

This Certificate is for the exclusive use of Intertek's Client and is provided pursuant to the Certification agreement between Intertek and its Client. Intertek's responsibility and liability are limited to the terms and conditions of the agreement. Intertek assumes no liability to any party, other than to the Client in accordance with the agreement, for any loss, expense or damage occasioned by the use of this certificate. Only the Client is authorized to permit copying or distribution of this certificate and then only in its entirety. Use of Intertek's Certification mark is restricted to the conditions laid out in the agreement. Any further use of the Intertek name for the sale or advertisement of the tested material, product or service must first be approved in writing by Intertek. Initial Factory Assessments and Follow up Services are for the purpose of assuring appropriate usage of the Certification mark in accordance with the agreement, they are not for the purposes of production quality control and do not relieve the Client of their obligations in this respect.





Reference No. 1917153

The 3rd "\*" denote the standard rated output voltage designation, which can be "12" to "54" or "12.0" to "54.0" in 0.1V increments

The 4th "\*"

- =-T2 means desktop class II with C8 AC inlet
- =-T2A means desktop class II with C18 AC inlet
- =-T3 means desktop class I
- =-T3TAB means desktop class I
- =-T3A means desktop class I
- =-TW means desktop class II with input wire without plug
- =-TW3 means desktop class I
- =-TP means desktop class II with power supply cord with plug
- =-TP3 means desktop class I
- =-P2 means Encapsulated Type, class II, with two-core input wire, IP68
- =-P3 means Encapsulated Type, class I, with two-core input wire, IP68

The last \* denote any six character = 0-9 or A-Z or ()[] or – or blank for marketing purposes.

The models which 4th"\*" is T2, T2A, TW,TP,P2 comply with EN60601-1: 2006 + A1 +A11 +A12 and EN 60601-1-11:2015; other models comply with EN60601-1:2006 + A1 +A11 +A12

## Attachment:

## GT\*41133-\*\*\*\* Desktop models and open frame models

Model	Rated output voltage range	Max. rated output current	Max. rated output power
GT*41133-*16***	12-16Vdc	7.5A	90W
GT*41133-*24***	16.1-24Vdc	5.6A	90W
GT*41133-*35***	24.1-35Vdc	3.73A	90W
GT*41133-*48***	35.1-48Vdc	2.56A	90W

GT\*96900P\*\*\*\* GT\*961200P\*\*\*\* Desktop models

Model	Rated output voltage range	Max. rated output current	Max. rated output power
GT*96900P****	12-54Vdc	7.5A	90W
GT*961200P****	12-14.9Vdc	9.2A	111W
	15-54Vdc	8A	120W

This Certificate is for the exclusive use of Intertek's Client and is provided pursuant to the Certification agreement between Intertek and its Client. Intertek's responsibility and liability are limited to the terms and conditions of the agreement. Intertek assumes no liability to any party, other than to the Client in accordance with the agreement, for any loss, expense or damage occasioned by the use of this certificate. Only the Client is authorized to permit copying or distribution of this certificate and then only in its entirety. Use of Intertek's Certification mark is restricted to the conditions laid out in the agreement. Any further use of the Intertek name for the sale or advertisement of the tested material, product or service must first be approved in writing by Intertek. Initial Factory Assessments and Follow up Services are for the purpose of assuring appropriate usage of the Certification mark in accordance with the agreement, they are not for the purposes of production quality control and do not relieve the Client of their obligations in this respect.

Page 3 of 4





Reference No. 1917153

Manufacturing site(s)

GlobTek (Suzhou) Co., Ltd Building 4, No. 76 JinLing East Road, Suzhou Industrial Park, Suzhou,

JiangSu, 215021, China

The instruction for use shall be written in a language acceptable according to the national regulation in the country where the product is to be used.

If the product is to be marketed in different European countries, it must be provided with a certified plug appropriate for each country.

This certificate replaces previously issued ref. No. 1713928 dated 10 July 2017. A new certificate has been issued on account of:

- 1. Updated the Model Similarity in the general product information
- 2. Updated the list of critical components in table 8.10
- 3. Added one photo "GT\*96900P series, GT\*961200P series External view"

This certificate replaces previous issued certificate with the same number, dated 2019-10-29, due to editorial corrections.

Certification Body

Intertek Semko AB, Product Certification

Place Kista - Stockholm

Signed

Henrik Wikström

Date

14 October 2019

This Certificate is for the exclusive use of Intertek's Client and is provided pursuant to the Certification agreement between Intertek and its Client. Intertek's responsibility and liability are limited to the terms and conditions of the agreement. Intertek assumes no liability to any party, other than to the Client in accordance with the agreement, for any loss, expense or damage occasioned by the use of this certificate. Only the Client is authorized to permit copying or distribution of this certificate and then only in its entirety. Use of Intertek's Certification mark is restricted to the conditions laid out in the agreement. Any further use of the Intertek name for the sale or advertisement of the tested material, product or service must first be approved in writing by Intertek. Initial Factory Assessments and Follow up Services are for the purpose of assuring appropriate usage of the Certification mark in accordance with the agreement, they are not for the purposes of production quality control and do not relieve the Client of their obligations in this respect.