



Ref. Certif. No.

SE-111607

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Medical Power Supply

Name and address of the applicant

GlobTek, Inc.
186 Veterans Drive Northvale NJ 07647, USA

Name and address of the manufacturer

Same as applicant

Name and address of the factory

Additional Information on page 2

Note: When more than one factory, please report on page 2

Ratings and principal characteristics

Input: 100-240V~, 50-60Hz, 0.3A;
Output:
GT*46101-*05*-USB: 5Vdc, 2A max.;
GT*46101-*06*-USB: 5.1-5.5Vdc, 2.54A max.

Trademark / Brand (if any)



Customer's Testing Facility (CTF) Stage used

-

Model / Type Ref.

GT*46101-*05*-USB, GT*46101-*06*-USB

Additional information (if necessary may also be reported on page 2)

Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005+A1+A2
IEC 60601-1-6:2010+A1+A2
IEC 60601-1-11:2015+A1

As shown in the Test Report Ref. No. which forms part of this Certificate

230500749SHA-001, 230500749SHA-002,
230500749SHA-003

This CB Test Certificate is issued by the National Certification Body

Intertek Semko AB
Torshamnsgatan 43
Box 1103
SE-164 22 Kista, Sweden

Date: 19 July, 2023

intertek

Signature:

Hyden L



Ref. Certif. No.

SE-111607

Factories

- 1. GlobTek, Inc.
186 Veterans Drive Northvale NJ 07647, USA
- 2. GlobTek (Suzhou) Co., Ltd
Building 4, No. 76, Jin Ling East Rd., Suzhou Industrial Park, Suzhou, JiangSu 215021, China
- 3. Shenzhen ENG Electronics Co., Ltd.
Block B, Nuclear Group Industrial District, Baishixia, Fuyun Town, Bao'an, Shenzhen, China

Additional information

Explanation of models GT*46101-*05*-USB, GT*46101-*06*-USB:
The 1st symbol "*" can be "M" or "-" or "H" for market identification and not related to safety.
The 2nd symbol "*" can be "01" to "13", with interval of 1, denote the rated output wattage designation.
The 3rd symbol "*" can be "-0.5" to "-0.9" with interval of 0.1, or blank indicate no voltage different, optional deviation, subtracted from standard output voltage.
The "05" or "06" and 3rd symbol "*" together denote the output voltage, with a range of 5-5.5 volts.

The national difference of USA, Canada have been checked.

Only plugs for EN 50075:1990(two-pole plug) and AS/NZS 3112:2017 have been evaluated in the report.

The following applicable requirements have not been evaluated by Intertek:
-Biocompatibility according to ISO 10993
-EMC according to IEC 60601-1-2

Date: 19 July, 2023

Signature: