

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME**CB TEST CERTIFICATE**

Product

Medical Power Supply

Name and address of the applicant

GlobTek, Inc.
186 Veterans Dr. Northvale, NJ 07647 USA

Name and address of the manufacturer

Same as applicant

Name and address of the factory

Note: When more than one factory, please report on page 2

See page 2

Ratings and principal characteristics

Input: 100-240V~, 50-60Hz or 50/60Hz, 2.2A;
Output: 12-54VDC, Max.13.33A, Max. 180W

Trademark (if any)



Customer's Testing Facility (CTF) Stage used

-

Model / Type Ref.

GT*961600P****, GT*961800P****

Additional information (if necessary may also be reported on page 2)

See page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005+A1
IEC 60601-1-6:2010+A1
IEC 60601-1-11:2015

As shown in the Test Report Ref. No. which forms part of this Certificate

210600863SHA-001, 210600863SHA-002, 210600863SHA-003

This CB Test Certificate is issued by the National Certification Body

Intertek Semko AB
Torshamnsgatan 43
Box 1103
SE-164 22 Kista, Sweden**intertek**

Signature:



Gary Hu

Date: 29 September, 2021

Factories

1. GlobTek, Inc.
186 Veterans Dr. Northvale, NJ 07647 USA
2. GlobTek (Suzhou) Co., Ltd
Building 4, No. 76 JinLing East Road, Suzhou
Industrial Park, Suzhou, 215021, JiangSu, China

Additional information

Explanation of models GT*961600P****, GT*961800P****

The 1st “*” part can be ‘M’ or ‘-’ or ‘H’ for market identification and not related to safety.

The 2nd “*” denotes the rated output wattage designation, which can be “01” to “180”, with interval of 1W

The 3rd “*” denote the standard rated output voltage designation, which can be “12” to “54” or “12.0” to “54.0” in 0.1V increments

The 4th “*”

- =-T2 means desktop class II with C8 AC inlet
- =-T2A means desktop class II with C18 AC inlet
- =-T3 means desktop class I or class II with functional earth with C14 AC inlet
- =-T3A means desktop class I or class II with functional earth with C6 AC inlet
- =-TW means desktop with input wires without plug
- =-TP means desktop with power cord and plug

The last * denote any six character = 0-9 or A-Z or ()[] or – or blank for marketing purposes.

The national differences for Canada, USA, Switzerland, Korea, Japan have been checked.

The following applicable requirements have not been evaluated by Intertek:

- Biocompatibility according to ISO 10993
- EMC according to IEC 60601-1-2

Date: 29 September, 2021

Signature:

