



Ref. Certif. No.

SE-103706

**IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR
ELECTRICAL EQUIPMENT (IECEE) CB SCHEME**

CB TEST CERTIFICATE

Product

Medical Power Supply

Name and address of the applicant

GlobTek, Inc.
186 Veterans Dr. Northvale, NJ 07647 USA

Name and address of the manufacturer

Same as applicant

Name and address of the factory

Note: When more than one factory, please report on page 2

See page 2

Ratings and principal characteristics

Input: 100-240V~, 50-60Hz or 50/60Hz, 3.0A;
Output: 12.0-54.0VDC, Max. 18.75A, Max. 225W.

Trademark (if any)



Customer's Testing Facility (CTF) Stage used

-

Model / Type Ref.

GT*96225*P*****_*

Additional information (if necessary may also be
reported on page 2)

See page 2

A sample of the product was tested and found
to be in conformity with

IEC 60601-1: 2005+A1
IEC 60601-1-6:2010+A1
IEC 60601-1-11:2015+A1

As shown in the Test Report Ref. No. which forms part
of this Certificate

200501730SHA-001, 200501730SHA-002,
200501730SHA-003

This CB Test Certificate is issued by the National Certification Body

Intertek Semko AB
Torshamnsgatan 43
Box 1103
SE-164 22 Kista, Sweden

intertek

Signature:

Gary Hu

Date: 20 January, 2021

Factories

GlobTek, Inc.
186 Veterans Dr. Northvale, NJ 07647 USA

GlobTek (Suzhou) Co., Ltd
Building 4, No. 76 JinLing East Road, Suzhou Industrial Park, Suzhou, JiangSu, 215021, China

Additional information

Explanation of model GT*96225*P*****-* :

The 1st "*" part can be 'M' or '-' or 'H' for market identification and not related to safety.

The 2nd "*" can be 0, 1, 2 or 3, denote the different mechanical construction, "0" means open frame, "1" means L frame, "2" means cage, "3" means potted.

The 3rd "*" can be "001" to "225", denotes the rated output wattage designation from 1W to 225W, in step of 1 denote 1W.

The 4th "*" can be "12" to "54" or "12.0" to "54.0", denote the standard rated output voltage designation from 12.0V to 54.0Vdc, in step of 0.1 denote 0.1V.

The 5th "*" can be optional, blank or A to H, denote the AUX Output voltage code.

The 6th "*" can be Blank, -C or -D, related to PCB size, Blank=2"x4", -C=3"x5", -D= 7"x4.22".

The 7th "*" =-F or F means Open Frame class I or class II with functional earth.

=-FW or FW means Open Frame class II.

=-P2 or P2 means Encapsulated Type, class II.

=-P3 or P3 means Encapsulated Type, class I or class II with functional earth.

The last * denote any six character, which can be 0-9 or A-Z or ()[] or - or blank for marketing purposes, -* can be blank.

The group and national differences for the CENELEC countries and the national differences for Canada, USA, Switzerland, Korea, Japan have been checked.

Date: 20 January, 2021

Signature: 